Comparison of two computational models for the inverse problem of electroencephalography[*]

Ana Alonso R.[†]   Jessika Camaño[‡]   Rodolfo Rodríguez[§]   Alberto Valli[¶]

Abstract

The goal of this work is to assess two approximation methods for the inverse problem of electroencephalography: the localization of brain activity from measurements of the electric potential on the surface of the head. The source current is modeled as a dipole whose localization and polarization has to be determined. The two considered methods are the so called *subtraction approach* and *direct approach*. The former is based on subtracting a fundamental solution, which has the same singular character of the actual solution, and solving computationally the resulting non-singular problem. Instead, the latter consists in solving directly the problem with singular data by means of an adaptive process based on an a posteriori error estimator, which allows creating meshes appropriately refined around the singularity. A set of experimental tests for both, the forward and the inverse problem, are reported. The main conclusion of these tests is that the direct approach combined with adaptivity is preferable when the localization of the dipole is close to an interface.

Key words: inverse problem, dipole source, electrostatic, electroencephalography

Mathematics subject classifications: 65N15, 65N21, 65N30

References

[1] Alonso Rodríguez, A., Camaño, J., Rodríguez, R. and Valli, A., *A posteriori error estimates for the problem of electrostatics with a dipole source.* preprint (2013).

[2] Valli, A., *Solving an electrostatics-like problem with a current dipole source by means of the duality method.* Appl. Math. Lett., vol. 25, pp. 1410-1414, (2012).

[3] WOLTERS, C. H., KÖSTLER, H., MÖLLER, C., HÄRDTLEIN, J., GRASEDYCK, L. AND HACKBUSCH, W., *Numerical mathematics of the subtraction method for the modeling of a current dipole in EEG source reconstruction using finite element head models.* SIAM J. Sci. Comput., vol. 30, pp. 24-45, (2007).